



TD2014NA

Summary of Major Modifications

The Technical Document on Harmonization of Analysis and Reporting of 19-Norsteroids Related to Nandrolone has undergone a brief revision by WADA's Laboratory Expert Group (LabEG). The new version of the document, TD2014NA, includes modifications to the interpretation and reporting of results following the application of the GC-C-IRMS method for 19-NA determinations:

- i). Application of the maximum allowed measurement uncertainty (u_{C-Max}) of 1.0 ‰ to $\delta^{13}C$ values (point iii- of footnote 1).
- ii). Rewording of test report: reporting of $\delta^{13}C$ values and associated u_c . It is no longer necessary to report the $\Delta\delta$ values (footnote 4).

These modifications are consistent with the new Technical Document on Detection of Synthetic Forms of Endogenous Anabolic Androgenic Steroids by GC-C-IRMS (TD2014IRMS).