

TD2019DL v. 2.0

Summary of Major Modifications

The Technical Document on Decision Limits for the Confirmatory Quantification of Threshold Substances, TD2019DL version 2.0, has been modified to include important clarifications regarding the permitted use of densitometers for the determination of urinary specific gravity (SG) during the Initial Testing Procedure(s), as well as to provide specific instructions on how to conduct and report SG measurements during the “A” and “B” Confirmation Procedures (see footnote 1).

- In addition, footnote a. under Table 1 clarifies that the Decision Limits have been rounded up to the second significant figure. When reporting quantification results for Threshold Substances, Laboratories shall report the mean value from triplicate determinations, rounded down (truncated) to the same number of **decimal places** as the applicable DL (see Art. 4.1).

The TD2019DL v. 2.0 replaces the former TD2019DL v 1.0.