

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA <u>Laboratory Expert Group</u>		
Date:	21 December 2020	Effective Date:	1 May 2021

Detection of Synthetic Forms of *Prohibited Substances* by GC/C/IRMS

1.0 Introduction

This *Technical Document (TD)* describes the Analytical Method to detect the presence of synthetic forms of *Prohibited Substances* by Gas Chromatography / Combustion / Isotope Ratio Mass Spectrometry (GC/C/IRMS) in urine *Samples*.

19-Norandrosterone (19-NA) and 19-noretiocholanolone (19-NE) are considered in a separate *TD* (TD 19NA ^[1]), and the technical recommendations and requirements described herein shall not be applied to their analysis.

1.1 Application of GC/C/IRMS

1.1.1 GC/C/IRMS Analysis as a Confirmation Procedure (CP) for the Administration of Synthetic Forms of Endogenous Anabolic Androgenic Steroids (EAAS)

GC/C/IRMS analysis shall be conducted as a CP for atypical steroid profiles, as described in the TD EAAS ^[2]. In addition, even if the *Markers* of the steroid profile are within the normal ranges, the Testing Authority, the Results Management Authority, or WADA may request GC/C/IRMS analysis to be performed on any urine *Sample*.

Furthermore, the Laboratory or the Athlete Passport Management Unit (APMU) may at any time advise the Testing Authority to perform (or not) the GC/C/IRMS analysis based upon its expertise, for example in the presence of any other *Marker(s)* of administration of EAAS.

1.1.2 GC/C/IRMS Analysis as a CP for the administration of Synthetic Forms of Other *Prohibited Substances*

GC/C/IRMS analysis shall be conducted before reporting an *Adverse Analytical Finding (AAF)* for the following compounds when their estimated concentrations (SG-adjusted, if needed) are determined as follows:

[Comment: When the measured SG of the urine Sample (SG_{Sample}) is greater than (>) 1.018, the concentrations (free and hydrolyzed glucuroconjugated steroids) shall be adjusted according to the Eq. 1:

$$(Eq. 1) \text{ Conc}_{adj} = \frac{(1.020 - 1)}{SG_{Sample_Max} - 1} \cdot \text{Conc}_{measured}$$

Refer to effective TD DL ^[3] for instructions on calculating SG_{Sample_Max}]

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA <u>Laboratory Expert Group</u>		
Date:	21 December 2020	Effective Date:	1 May 2021

- *Samples* containing Boldenone (B) or the main Boldenone *Metabolite* (BM1; 5β-androst-1-en-17β-ol-3-one) at a concentration estimated between (≥) 2.5 ng/mL and (≤) 30 ng/mL:

- Confirmed findings (following the application of a CP confirming the identity of the Analyte in accordance with the TD IDCR ^[4]) for B and/or BM1 at concentrations estimated below (<) 2.5 ng/mL should be reported as *Atypical Findings (ATF)*, unless the results of the GC/C/IRMS analysis, if performed (depending on Laboratory's analytical capacity and following consultation with the Testing Authority), conclusively establish the exogenous (AAF) or endogenous (Negative Finding) origin of the substance;
- Confirmed findings for B and/or BM1 at a concentration estimated above (>) 30 ng/mL should be reported as AAF without the need for GC/C/IRMS analysis, unless the *Sample* shows signs of extensive degradation ^[2], in which case the result shall be reported as an ATF.

[Comment: the GC-MS identification of the Analytes (target compounds and endogenous reference compounds) is also required for reporting an AAF or ATF based on the GC/C/IRMS analysis. See Article 2.2.]

- *Samples* containing 6α-hydroxy-androstenedione (6α-OH-AD) at a concentration estimated above (>) 10 ng/mL with no signs of extensive *Sample* degradation (see TL21 ^[5] and TD EAAS ^[2]):

- Findings for 6α-OH-AD at a concentration estimated below or at (≤) 10 ng/mL or above (>) 10 ng/mL and in the presence of signs of extensive *Sample* degradation ^[2] should be reported as a Negative Finding unless the Laboratory, at its discretion, decides to proceed with the GC/C/IRMS analysis (in which case the results of the analysis shall be reported in accordance with this TD).

*[Comment: 6α-OH-AD is also a Marker of the administration of EAAS (e.g. testosterone, DHEA, androstenedione). Therefore, it is recommended that for *Samples* with a 6α-OH-AD concentration > 10 ng/mL and no signs of extensive *Sample* degradation, the Laboratory, in consultation with the Testing Authority, performs GC/C/IRMS analysis targeting the Markers of the steroid profile before determining the need to subject the *Sample* to GC/C/IRMS analysis for 6α-OH-AD (e.g. if the GC/C/IRMS results for the Markers of the steroid profile are negative) . This is particularly important for those Laboratories that do not have the analytical capacity to do the test for 6α-OH-AD and would have to subcontract the analysis to another Laboratory.*

Nevertheless, the Laboratory may also proceed directly to the GC/C/IRMS analysis for 6α-OH-AD if there is in-house analytical capacity to perform the analysis and/or if the Laboratory has an agreement with the Testing Authority to subcontract the GC/C/IRMS analysis at the Laboratory's discretion.]

- *Samples* containing Formestane (F) at a concentration estimated between (≥) 50 ng/mL and (≤) 150 ng/mL:

- Findings for F at a concentration estimated below (<) 50 ng/mL should be reported as a Negative Finding, unless the result of the GC/C/IRMS analysis, if performed (depending on Laboratory's analytical capacity and following consultation with the Testing Authority), conclusively establishes the exogenous origin of the substance (AAF);
- Confirmed findings for F at a concentration estimated above (>) 150 ng/mL shall be reported as AAF without the need to conduct a GC/C/IRMS analysis.

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA <u>Laboratory Expert Group</u>		
Date:	21 December 2020	Effective Date:	1 May 2021

- *Samples* containing Prednisone (PS) or Prednisolone (PSL) at a concentration estimated between (\geq) 30 ng/mL and (\leq) 60 ng/mL (see TL19 ^[6]):

- Findings for PS or PSL at a concentration estimated below ($<$) 30 ng/mL should be reported as a Negative Finding without the need for GC/C/IRMS analysis;
- Confirmed findings for PS or PSL at a concentration estimated above ($>$) 60 ng/mL shall be reported as AAF without the need for GC/C/IRMS analysis, unless the *Sample* shows signs of extensive degradation ^[2], in which case the *Sample* shall be reported as an ATF.

[Comment: Laboratories that do not have in-house analytical capacity to test for B, BM1, 6 α -OH-AD, F, PS or PSL shall, in consultation with the Testing Authority (or as previously agreed), subcontract the analysis to another Laboratory that has such analytical capability. In such cases, the Laboratory shall confirm the finding (in accordance with the TD IDCR ^[4]) before sending the Sample for GC/C/IRMS analysis to the other Laboratory.]

1.1.3 “B” Sample CP

When an AAF is reported based on the results of a GC/C/IRMS analysis performed on the “A” *Sample*, only the GC/C/IRMS analysis and GC-MS identification of the Analytes shall be repeated during the “B” *Sample CP*, if applicable (see Article 2.2.).

2.0 GC/C/IRMS Analysis

The application of GC/C/IRMS is based on the following:

- The determination of the $\delta^{13}\text{C}$ value of the Target Compound(s) (TCs), *i.e.* the *Marker(s)* of the urinary “steroid profile” [e.g. androsterone (A), etiocholanolone (Etio), 5 α -androstane-3 α ,17 β -diol (5 α Adiol), 5 β -androstane-3 α ,17 β -diol (5 β Adiol), testosterone (T), epitestosterone (E)], as well as other target Analytes [e.g. B, BM1, 6 α -OH-AD, F, PS, PSL or epiandrosterone (EpiA)];

[Comment: EpiA, which is detected in the steroid sulfoconjugated fraction, is recommended as an additional TC for GC/C/IRMS analysis to determine the administration of testosterone or its precursors. The Laboratory should consider performing GC/C/IRMS analysis for EpiA when the results obtained for the Markers of the steroid profile in the glucuroconjugated fraction are suspicious but inconclusive.]

- The determination of the $\delta^{13}\text{C}$ value of the Endogenous Reference Compound(s) (ERC), *i.e.* pregnanediol (PD), pregnanetriol (PT), 5 α -androst-16-en-3 α -ol (16-en), 11 β -hydroxy-androsterone (11-OH-A), and 11-oxo-etiocholanolone (11-oxo-Etio); and

[Comment: The Laboratory shall validate the use of, at least, PD as the principal ERC and one additional ERC from this list. See Articles 2.1.1 and 2.4.]

- The calculation of the absolute difference in $\delta^{13}\text{C}$ values, *i.e.* the $|\Delta\delta^{13}\text{C}|$ value, between the ERC(s) and the TC(s):

$$\text{(Eq. 2) } |\Delta\delta^{13}\text{C}| = |\delta^{13}\text{C}_{\text{ERC}} - \delta^{13}\text{C}_{\text{TC}}|$$

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA Laboratory Expert Group		
Date:	21 December 2020	Effective Date:	1 May 2021

2.1 GC/C/IRMS Test Method Characteristics

2.1.1 GC/C/IRMS Test Method Validation Requirements

As part of the Analytical Method validation, the Laboratory shall determine:

- Linearity of the Ion Source

The range of intensities of CO₂ pulses (in mV or nA, as applicable) that produce consistent results within the specifications of the instrument manufacturer, as determined from 6 to 10 pulses of different peak heights/intensities. The linearity of the ion source shall be checked regularly, e.g. monthly.

- Linearity of the Instrument

The range of intensities [expressed in response units (mV or nA) and/or as amounts (ng)] of injected Analytes that give a consistent and repeatable $\delta^{13}\text{C}$ value for standards of TC(s) and ERC(s), as determined after the replicate injection and analysis ($n \geq 3$) of 6 to 10 different dilutions of underivatized or acetylated RM(s).

The linear range of the instrument shall comply with the following criteria:

- The means of the replicate ($n \geq 3$) $\delta^{13}\text{C}$ determinations at each dilution level shall not differ by more than 0.5 ‰ from the mean $\delta^{13}\text{C}$ value of all $\delta^{13}\text{C}$ determinations within the linear range; and
- The standard deviation (SD) of the replicate ($n \geq 3$) $\delta^{13}\text{C}$ determinations at each dilution level shall not be higher than (<) 0.5 ‰.

- Limit of Quantification (LOQ) in Urine

The lowest concentration (ng/mL) of each TC and ERC in urine, when using not more than (\leq) 25 mL of Sample, that produces a measurable signal in the linear range of the instrument with a SD ($n \geq 3$), which is not higher than the corresponding u_c (i.e. $\text{SD} \leq u_c$), as determined by the Laboratory during method validation.

[Comment: In routine analysis, more than 25 mL of urine could be used at the Laboratory's discretion if the Sample volume is sufficient.]

- Measurement Uncertainty (MU)

The LOQ in urine and the maximum acceptable combined standard uncertainty (u_{c_Max}) for the determination of the $\delta^{13}\text{C}$ values of TCs and ERCs shall be not higher than (\leq) the values specified in Table 1 below.

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA Laboratory Expert Group		
Date:	21 December 2020	Effective Date:	1 May 2021

Table 1 LOQ and u_{c_Max} Requirements for TCs and ERCs

<u>Analyte</u>	<u>LOQ</u> (ng/mL)	u_{c_Max} (δ) (‰)
ERCs	50	0.7
T, 5αAdiol	10	0.7
5βAdiol	20	0.7
EpiA	20	1.0
E, F	50	1.0
A, Etio	100	0.7
B / BM1	2.5	1.0
6α-OH-AD	10	1.0
PS / PSL	30	1.0

- Reference Population Data

The $\Delta\delta^{13}C$ values for each ERC-TC pair [including as TCs, at least, A, Etio, 5 α Adiol, 5 β Adiol and T, and using at least two (2) ERCs, one of which shall be PD] analyzed in volunteer urine samples and/or *Athlete* negative *Samples* (a minimum of 20 male and 20 female urine samples), covering the range of steroid concentrations usually found in *Samples*.

[Comment: The requirement to have measurements from a minimum of 20 urine reference samples from females does not apply to T and E determinations. However, a minimum total number of 40 reference determinations is required.]

The analysis of these negative urine samples shall serve to evaluate the Fitness-for-Purpose of the GC/C/IRMS Analytical Method and shall meet the following criteria:

- The Mean $_{|\Delta\delta^{13}C|} + 2SD_{|\Delta\delta^{13}C|}$ value for ERC-TC combinations containing A shall be ≤ 2 ‰;
- The Mean $_{|\Delta\delta^{13}C|} + 2SD_{|\Delta\delta^{13}C|}$ value for ERC-TC combinations containing Etio, T or the Adiol as TCs shall be ≤ 3 ‰;
- The Mean $_{|\Delta\delta^{13}C|} + 2SD_{|\Delta\delta^{13}C|}$ value for ERC-TC combinations containing E or EpiA as TCs shall be ≤ 4.5 ‰ or ≤ 4 ‰, respectively;
- For ERC-TC combinations including A, Etio, 5 α Adiol, 5 β Adiol or T, the SD of all $|\Delta\delta^{13}C|$ values shall be ≤ 1.0 ‰; and
- For ERC-TC combinations including E and EpiA the SD of all $|\Delta\delta^{13}C|$ values shall be ≤ 1.2 ‰.

- Endogenous Reference Compounds (ERCs)

The Laboratory shall validate the use of at least two (2) ERCs, including PD as the primary ERC (ERC₁), and at least one (1) additional ERC selected from 16-en, 11-OH-A, 11-oxo-Etio and PT.

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA <u>Laboratory Expert Group</u>		
Date:	21 December 2020	Effective Date:	1 May 2021

- Target Compounds (TCs)

The Laboratory shall validate the analysis of, at least, the *Markers* of the urinary steroid profile (A, Etio, T, 5 α Adiol and 5 β Adiol).

2.1.2 GC/C/IRMS Analysis Requirements

The Laboratory shall implement the following into the GC/C/IRMS methodology:

- The GC/C/IRMS analysis shall be conducted on a single *Sample Aliquot*.
- System Calibration – Use of Certified Reference Materials (CRM).

The system shall be calibrated periodically (e.g. quarterly) against a steroid CRM containing a mixture of steroids [e.g. CU/PCC 33-2, 34-3, 44-1, MX018-1, MX018-2, MX018-3 or other mixture of certified steroid(s)] with $\delta^{13}\text{C}$ values that are traceable to the assigned values of the recognized international Reference Material (RM). Major revisions of the system (e.g. change of reference gas, cleaning of the ion source, change of oxidation furnace) shall require a check of the calibration of the system;

[Comment: The use of a mixture of certified steroids covering the range of $\delta^{13}\text{C}$ values that may be found for TCs and ERCs in urine (e.g. -17 ‰ to -34 ‰) is recommended.]

- Stability of CO₂ Pulses

The stability of CO₂ pulses (precision of the 45/44 and 46/44 ratios) shall be tested before the analysis of each batch of *Samples* and shall meet manufacturer's specifications.

- *Sample Preparation*

The urinary TC(s) and ERC(s) once hydrolyzed shall be further purified by High Performance Liquid Chromatography (HPLC) or similarly Fit-for-Purpose Sample preparation technique (e.g. multidimensional gas chromatography) prior to the GC/C/IRMS analysis.

- Derivatization

The steroids may be analyzed underivatized or after acetylation, but only values equivalent to underivatized compounds shall be used to determine the $|\Delta\delta^{13}\text{C}|$ value of the ERC-TC pair. The following mass balance equation for adjustment of the measured $\delta^{13}\text{C}$ values from acetates back to the free form shall be used:

$$\text{(Eq. 3)} \quad \delta^{13}\text{C}_s = (n_{cd} \cdot \delta^{13}\text{C}_{cd} - n_d \cdot \delta^{13}\text{C}_{d_corr}) / n_s$$

where C: concentration; n: number of carbon atoms; s: native steroid (underivatized form); d: derivative group (e.g. acetyl), and cd: derivatized compound. As $\delta^{13}\text{C}_d$ is not known, $\delta^{13}\text{C}_{d_corr}$ is estimated empirically by consecutive measurements of a non-acetylated and acetylated steroid (e.g. 16-en or 5 α -androstanol).

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA Laboratory Expert Group		
Date:	21 December 2020	Effective Date:	1 May 2021

- Use of Reference Materials (RM)

The Laboratory shall implement appropriate CRM or RM (with traceable $\delta^{13}\text{C}$ values, which have been corrected against a CRM) of the relevant free (unconjugated) TC(s) and ERC(s), underivatized or acetylated, as appropriate. The solution(s) of CRM / RM shall be injected at the beginning and at the end of each sequence of analysis (e.g. after 25-30 injections). The consistency of the $\delta^{13}\text{C}$ determinations of the CRM / RM shall be monitored through the use of RM control charts.

- Use of Quality Control (QC) Samples

A negative (NQC) and a positive (PQC) urine quality control sample containing all relevant TC(s) and ERC(s) shall be included in each batch of *Samples* analyzed and subjected to the same *Sample* preparation procedure (including hydrolysis of conjugates and Analyte derivatization, if applicable).

- NQC: $\delta^{13}\text{C}$ values of TCs and ERCs in a normal endogenous range (i.e. between -16 ‰ and -25 ‰), with $|\Delta\delta^{13}\text{C}|$ between ERC(s) and TC(s) not greater than (\leq) the applicable decision rule(s) for a GC/C/IRMS positive test specified in Article 2.3.2 and Annex B;
- PQC: $\delta^{13}\text{C}$ value of ERC in a normal endogenous range (i.e. between -16 ‰ and -25 ‰), with a $|\Delta\delta^{13}\text{C}|$ between ERC(s) and relevant TC(s) greater than ($>$) the applicable decision rule(s) for GC/C/IRMS positive test specified in Article 2.3.2 and Annex B.

The consistency of the $\delta^{13}\text{C}$ determinations of the NQC and PQC shall be monitored through the use of QC-charts.

[Comment: An NQC is not required for GC/C/IRMS determinations of B or BM1, 6 α -OH-AD, F, PS or PSL.]

- Use of Endogenous Reference Compounds (ERCs)

- The Laboratory shall use two (2) ERCs (including PD as the ERC₁, if measurable in the *Sample*) to conclude an AAF (see Article 2.4.2);
- The same ERCs shall be used for determination of all $|\Delta\delta^{13}\text{C}|$ values in the *Samples* and in the QC (NQC, PQC) samples.

[Comment: 11-OH-A and 11-oxo-Etio shall not be considered together since they can be related to the same precursor (adrenosterone).]

- Target Compounds (TC)

The Laboratory shall be capable of measuring the $\delta^{13}\text{C}$ -values of, at least, A, Etio, T, 5 α Adiol and 5 β Adiol. When the concentration is sufficient, the TC(s) should be selected/prioritized depending on the *Marker(s)* of the “steroid profile” that prompted the GC/C/IRMS analysis:

- T, 5 α Adiol and 5 β Adiol are the main TC(s) to detect the administration of T;

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA <u>Laboratory Expert Group</u>		
Date:	21 December 2020	Effective Date:	1 May 2021

- The Laboratory shall perform GC/C/IRMS analyses for E on those *Samples* in which the concentration of E is abnormally high, *i.e.* if > 200 ng/mL in males or > 50 ng/mL in females (SG-adjusted, if needed);

[Comment: However, GC/C/IRMS analyses for E may also be performed at lower concentrations, at the Laboratory's discretion. A measured lower concentration of E does not invalidate an AAF or ATF resulting from the GC/C/IRMS analysis.]

- The Laboratory should consider performing GC/C/IRMS analysis for EpiA when the results obtained for the *Markers* of the steroid profile in the glucuroconjugated fraction are suspicious but inconclusive (*i.e.* when the criteria for reporting an AAF are not met but, in the Laboratory's opinion, the results are not consistent with the endogenous origin of the urinary *Markers*).

2.2 Identification of TC(s) and ERC(s) prior to reporting an AAF or ATF

The same *Sample Aliquot(s)* that were subjected to GC/C/IRMS analysis shall be analyzed by GC-MS under similar chromatographic conditions to ensure the identity of the peaks of the relevant TC(s) and ERC(s) and the absence of significant interference prior to reporting an AAF or an ATF based on GC/C/IRMS results.

Minor differences in retention times (RT) between the two techniques are expected. The provisions of the TD IDCR^[4] shall be followed. In addition, a full scan spectrum shall be obtained over the complete width of the steroid chromatographic peak(s) of interest to document the lack of interference.

GC-MS identification is not necessary when the GC/C/IRMS results are negative.

2.3 Interpretation of GC/C/IRMS results

No value obtained from any peaks of intensity below or above the validated linearity range of the instrument or in the presence of significant co-eluting peaks shall be considered or reported.

The results of the GC/C/IRMS analyses, obtained on the basis of each ERC used, shall be interpreted as follows:

2.3.1 Negative Result

When $|\Delta\delta^{13}\text{C}|$ values do not confirm the exogenous origin of the TC(s), *i.e.* when the $|\Delta\delta^{13}\text{C}|$ values of the TC-ERC pairs do not meet any of the criteria specified in Article 2.3.2.

2.3.2 Positive Result

When $|\Delta\delta^{13}\text{C}|$ value(s) are consistent with the exogenous origin of the TC(s), *i.e.* if one of the following sets of criteria is fulfilled (see Annex B):

- i. The $|\Delta\delta^{13}\text{C}|$ values of the ERC-T pair and of one of the ERC-5 α Adiol or ERC-5 β Adiol pairs are both > 3.0 ‰;
- ii. The $|\Delta\delta^{13}\text{C}|$ values of the ERC-5 α Adiol and ERC-5 β Adiol pairs are both > 3.0 ‰;

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA <u>Laboratory Expert Group</u>		
Date:	21 December 2020	Effective Date:	1 May 2021

- iii. The $|\Delta\delta^{13}\text{C}|$ value of the ERC-E pair is $> 4.5\text{‰}$;
- iv. The $|\Delta\delta^{13}\text{C}|$ value of the ERC-A pair is $> 2.0\text{‰}$ and the $|\Delta\delta^{13}\text{C}|$ value of the ERC-Etio pair is $> 3.0\text{‰}$, whereas T, 5 α Adiol and 5 β Adiol cannot be measured in the *Sample* due to, for example, low concentrations or peak interferences;

[Comment: This criterion for the ERC-A and ERC-Etio pairs shall not be applied alone if T, 5 α Adiol and 5 β Adiol are measurable in the Sample].

- v. The $|\Delta\delta^{13}\text{C}|$ value of the ERC-A pair is $> 2.0\text{‰}$ or the $|\Delta\delta^{13}\text{C}|$ value of the ERC-Etio pair is $> 3.0\text{‰}$, and the $|\Delta\delta^{13}\text{C}|$ value of one of the ERC-5 α Adiol or ERC-5 β Adiol pairs is $> 3.0\text{‰}$;
- vi. The $|\Delta\delta^{13}\text{C}|$ value of the ERC-5 α Adiol pair is $> 4.0\text{‰}$ and the $\delta^{13}\text{C}$ value of 5 α Adiol is $\leq -27.0\text{‰}$ (e.g. DHT administration);
- vii. The $|\Delta\delta^{13}\text{C}|$ value of either the ERC-B, ERC-BM1, ERC-F, ERC-6 α -OH-AD, ERC-EpiA, ERC-PS or ERC-PSL pair is $> 4.0\text{‰}$;

[Comment:

- *It is not expected that all TCs will be affected to the same extent;*
- *Decisions based on the $\Delta\delta^{13}\text{C}$ criteria specified in Article 2.3.2 i) to vii) and Annex B take into account the MU associated with the contributing $\delta^{13}\text{C}$ values;*
- *The Laboratory shall consider all analytical evidence (e.g. concentrations of “steroid profile” Markers, $|\Delta\delta^{13}\text{C}|$ values of other ERC-TC combinations, including the use of additional ERC(s) if necessary) in order to properly assess whether there is significant data or information that would cast doubt on or refute the reporting of the finding as an AAF or ATF.]*

2.3.3 Inconclusive Result

- When only one of the combined criteria specified in points i), ii), iv), v), or vi) above is met (e.g. the $|\Delta\delta^{13}\text{C}|$ value for the ERC-T pair is $> 3.0\text{‰}$ but the $|\Delta\delta^{13}\text{C}|$ values for the ERC-Adiol pairs are both $< 3.0\text{‰}$); or
- Due to technical limitations, e.g. when there is insufficient *Sample* volume or very low concentrations of TC(s) or ERC(s), or in the presence of interfering compounds or any other factor preventing a reliable measurement of the relevant diagnostic *Metabolite* or ERC-TC pair; or

[Comment: Technical limitations in the analysis of specific ERC-TC pair(s) should not necessarily invalidate the reporting of an AAF based on the reliable measurement of other ERC-TC pair(s) in the Sample and consideration of all available analytical evidence for interpretation of results.]

- The Laboratory may interpret the results as inconclusive when the criteria for reporting an AAF are not met but, in its opinion, are neither consistent with the endogenous origin of the urinary *Metabolites* (e.g. ERC $\delta^{13}\text{C}$ value at -25.0‰ and TC at -27.5‰).

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA <u>Laboratory Expert Group</u>		
Date:	21 December 2020	Effective Date:	1 May 2021

2.4 Conclusion of GC/C/IRMS Findings

2.4.1 Negative Finding

The Laboratory shall conclude a Negative Finding when the results of the analysis are negative (see Article 2.3.1), based on the use of PD as ERC₁, if measurable in the *Sample* (i.e. if the signal is not suppressed, affected by poor chromatography or by the administration of a precursor, e.g. pregnenolone). However, an additional validated ERC₂ shall be used to replace PD as ERC₁ when PD is not measurable in the *Sample*.

[Comment: When assessing a negative result based on the use of PD as ERC₁, or of an alternative ERC when PD is not measurable in the Sample, the Laboratory shall be cautious in the interpretation of the results when the $\delta^{13}\text{C}$ value of the ERC₁ seems not to be compatible with an endogenous origin (e.g. $\delta^{13}\text{C}$ (PD) ≤ -25.0 ‰). In such cases, the Laboratory shall evaluate a second ERC to assess compatibility between ERCs and consistency of GC/C/IRMS results (see below and Annex A).]

2.4.2 AAF

The Laboratory shall conclude an AAF based on the use of two (2) ERCs (including PD as the ERC₁, if measurable in the *Sample*) (see Annex A):

[Comment: When PD is not measurable in the Sample, or when in the Laboratory's opinion a negative or inconclusive result has been caused by a $\delta^{13}\text{C}$ (ERC₁) value that is not consistent with an endogenous origin (e.g. $\delta^{13}\text{C}$ (ERC₁) ≤ -25.0 ‰), the Laboratory shall use other two (2) ERCs (ERC₂ and ERC₃) to conclude an AAF (this is not necessary for a positive result when using PD as ERC₁; in such cases, the use of only one additional ERC is necessary for concluding an AAF) .

The Laboratory shall specify in its SOP which ERCs are designated as ERC₂ and ERC₃ (if applicable) and apply them consistently in this order for the analysis of Samples.

The validation of a third ERC is not a mandatory requirement for Laboratory compliance with this TD. If the analysis of an ERC₃ is needed, and the Laboratory has not validated a third ERC, then the Laboratory, in consultation with the Testing Authority (or Results Management Authority, if different), shall have the Sample sent for GC/C/IRMS analysis to another Laboratory that has a third validated ERC. If there is not enough Sample volume for a repeat GC/C/IRMS analysis of the "A" Sample using an ERC₃, a split "B" analysis should be considered and, if not possible, the result shall be reported as an ATF.]

The Laboratory shall conclude an AAF if the results are consistent between the two (2) ERCs (ERC₁ and ERC₂), i.e. if the results are positive (see Article 2.3.2) based on both ERCs.

[Example:

The $\delta^{13}\text{C}$ values of T and 5 β Adiol are measured in a *Sample* at -27.9 ‰ and -27.5 ‰, respectively. These results meet the combined positivity criterion for these TCs (see Article 2.3.2) when either PD or 16-en are used as ERC:

- ERC₁ = PD

$$\delta^{13}\text{C} (\text{PD}) = -24.3 \text{ ‰}; \quad |\Delta\delta^{13}\text{C}| (\text{PD-T}) = 3.6 \text{ ‰}; \quad |\Delta\delta^{13}\text{C}| (\text{PD-5}\beta\text{Adiol}) = 3.2 \text{ ‰}.$$

- ERC₂ = 16-en

$$\delta^{13}\text{C} (16\text{-en}) = -23.9 \text{ ‰}; \quad |\Delta\delta^{13}\text{C}| (16\text{en-T}) = 4.0 \text{ ‰}; \quad |\Delta\delta^{13}\text{C}| (16\text{en-5}\beta\text{Adiol}) = 3.6 \text{ ‰}.$$

Therefore, the Laboratory shall report the finding as an AAF for T and 5 β Adiol.]

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA <u>Laboratory Expert Group</u>		
Date:	21 December 2020	Effective Date:	1 May 2021

2.4.3 ATF

The Laboratory shall conclude an *ATF* when:

- The results of the analysis using the ERC₁ (PD or a substitute ERC, if PD is not measurable in the *Sample*) are inconclusive (see Article 2.3.3);
- If, when using an ERC₂ to verify the positive results obtained with the ERC₁, the GC/C/IRMS results obtained with the two (2) ERCs are not consistent between them, e.g. if the results are positive (see Article 2.3.2) based on the ERC₁, but negative or inconclusive when using the ERC₂.

[Example:

The $\delta^{13}\text{C}$ values of T and 5 β Adiol are measured in a *Sample* at -27.9 ‰ and -27.5 ‰, respectively. These results meet the combined positivity criterion for these two TCs (see Article 2.3.2) when PD is used as ERC₁ but fail to meet this criterion when the Laboratory applies 16-en as a second validated ERC₂:

- ERC₁ = PD
 $\delta^{13}\text{C}(\text{PD}) = -24.3 \text{ ‰}$; $|\Delta\delta^{13}\text{C}|(\text{PD-T}) = 3.6 \text{ ‰}$; $|\Delta\delta^{13}\text{C}|(\text{PD-5}\beta\text{Adiol}) = 3.2 \text{ ‰}$.
- ERC₂ = 16-en
 $\delta^{13}\text{C}(16\text{-en}) = -25.3 \text{ ‰}$; $|\Delta\delta^{13}\text{C}|(16\text{en-T}) = 2.6 \text{ ‰}$; $|\Delta\delta^{13}\text{C}|(16\text{en-5}\beta\text{Adiol}) = 2.2 \text{ ‰}$.

Since the results with the two (2) ERCs are not consistent, these results are inconclusive. Therefore, the Laboratory shall report the finding as an *ATF* for T and 5 β Adiol.]

- If there is not enough *Sample* volume for a repeat GC/C/IRMS analysis of the *Sample* using an ERC₃ (if needed).

2.5 Further Interpretation of GC/C/IRMS and GC-MSⁿ Results

The GC/C/IRMS and GC-MSⁿ CPs provide independent and complementary information, but their results must be considered together to arrive at a conclusion that is supported by the scientific literature and knowledge:

- The urinary steroid profile may show no major anomaly whilst being excreted following the administration of an EAAS; in such a case, the results of the GC/C/IRMS analysis indicating a synthetic origin of the steroid *Marker(s)* shall prevail. However, it is recommended that in such cases the Laboratory seeks a second opinion, in writing, from another Laboratory before reporting the *AAF*. The second opinion shall be recorded in the Laboratory Documentation Package;
- Conversely, values for *Marker(s)* of the “steroid profile” may be outside the subject-based longitudinal reference range while still being of endogenous origin (e.g. heavy ethanol drinking leading to an increased urinary excretion of T and 5 β Adiol, microbial formation of free T, or intense, prolonged exercise increasing the excretion of A);
- The “steroid profile” may be altered by the administration of a synthetic preparation of a steroid related to testosterone having a relatively enriched $\delta^{13}\text{C}$ value, which could be close to the $\delta^{13}\text{C}$ values of endogenous urinary steroids and therefore may not be detected by GC/C/IRMS;

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA <u>Laboratory Expert Group</u>		
Date:	21 December 2020	Effective Date:	1 May 2021

- Similarly, synthetic preparations of other *Prohibited Substances* may also have enriched $\delta^{13}\text{C}$ values, which would produce negative GC/C/IRMS results.

3.0 Reporting GC/C/IRMS Results

The Laboratory shall report the results of the GC/C/IRMS analyses for each *Sample* individually in *ADAMS* as follows:

- The Test Report shall include the $\delta^{13}\text{C}$ value (‰) of all the TC(s) and the ERC(s) measured, and each associated u_c (expressed in ‰ to one decimal place). For the application of the decision rules (see Article 2.3 and Annex B), the $|\Delta\delta^{13}\text{C}|$ (‰) values are automatically calculated in *ADAMS* and expressed truncated to one decimal place.

[Comment: In cases of GC/C/IRMS analysis for the Markers of the Steroid Profile, the Laboratory shall explain in the Comments section of the Test Report those circumstances when T, 5 α Adiol or 5 β Adiol could not be measured in the Sample (for example, due to low concentrations or interferences.)]

*[Comment: When a result is reported as a Negative Finding or ATF based on the use of the primary ERC (e.g. PD if measurable in the Sample and consistent with an endogenous origin), only the $\delta^{13}\text{C}$ value of the ERC₁ shall be reported in *ADAMS*. However, when more than one ERC is needed to conclude an AAF or ATF, the $\delta^{13}\text{C}$ values of both ERCs measured shall be reported.*

If PD is not used as ERC₁, the Laboratory shall explain in the Comments section of the Test Report why PD was not used, e.g. PD not measurable in the Sample, PD $\delta^{13}\text{C}$ value not consistent with endogenous origin].

- Report the GC/C/IRMS finding as either:
 - Negative Finding: GC/C/IRMS results do not indicate an exogenous administration of the substance(s); or
 - *AAF*: GC/C/IRMS results are consistent with the exogenous administration of the substance(s), specifying the identity of the relevant TC(s) that produced a positive GC/C/IRMS finding; or
 - *ATF*: GC/C/IRMS results are inconclusive, specifying the identity of each relevant TC(s) that produced an inconclusive GC/C/IRMS finding.

[Examples for reporting GC/C/IRMS findings in the Test Report:

- Negative Finding: GC/C/IRMS results do not confirm the exogenous origin of steroids;
- *AAF*: GC/C/IRMS results are consistent with the exogenous origin of T and 5 β Adiol;
- *ATF*: GC/C/IRMS results for T and the Adiols are inconclusive.]

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA <u>Laboratory Expert Group</u>		
Date:	21 December 2020	Effective Date:	1 May 2021

3.1 Second Opinion for GC/C/IRMS

At its discretion, the Laboratory may seek a second opinion from one of the GC/C/IRMS Experts before a GC/C/IRMS AAF or ATF is reported in ADAMS.

[Comment: The List of GC/C/IRMS Experts that may provide second opinions on Laboratory GC/C/IRMS findings is published on WADA’s website and it may be modified or updated at any time, as determined by WADA:

<https://www.wada-ama.org/en/resources/governance/list-of-gccirms-experts>.)]

The Laboratory shall provide appropriate and sufficient CP analytical data, in accordance with the requirements established in Annex B of the TD LDOC ^[7], in order for the Expert to produce a second opinion. The Laboratory shall also produce any additional information or data (e.g. raw data) that the Expert may consider necessary for the second opinion review.

In those situations, the written conclusion of the second opinion provider shall be inserted as part of the Laboratory record in the Laboratory Documentation Package.

4.0 References


- [1] WADA Technical Document TD NA: Harmonization of Analysis and Reporting of 19-Norsteroids related to Nandrolone.
- [2] WADA Technical Document TD EAAS: Measurement and Reporting of Endogenous Anabolic Androgenous Steroid (EAAS) Markers of the Urinary Steroid Profile.
- [3] WADA Technical Document TD DL: Decision Limits for the Confirmatory Quantification of Exogenous Threshold Substances by Chromatography-based Analytical Methods.
- [4] WADA Technical Document TD IDCR: Minimum Criteria for Chromatographic-Mass Spectrometric Confirmation of the Identity of Analytes for Doping Control Purposes.
- [5] WADA Technical Letter TL21: *In situ* Formation of 4-androstene-3,6,17-trione (6-oxo) and *Metabolites*.
- [6] WADA Technical Letter TL19: *In situ* Formation of Prednisone and Prednisolone.
- [7] WADA Technical Document TD LDOC: Laboratory Documentation Packages.

[Current versions of WADA Technical Documents and Technical Letters may be found at <https://www.wada-ama.org/en/what-we-do/science-medical/laboratories>]

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA Laboratory Expert Group		
Date:	21 December 2020	Effective Date:	1 May 2021

Annex A. Selection and Use of ERCs



* PD shall be used as the primary ERC, if measurable in the *Sample*.

** If the Laboratory has not validated a third ERC, the Laboratory should, in consultation with the Testing Authority send the *Sample* for analysis to another Laboratory that has a third validated ERC.

WADA Technical Document – TD2021IRMS

Document Number:	TD2021IRMS	Version Number:	1.0
Written by:	WADA Science / IRMS Working Group	Approved by:	WADA Executive Committee
Reviewed by:	WADA Laboratory Expert Group		
Date:	21 December 2020	Effective Date:	1 May 2021

Annex B. Decision Rules for GC/C/IRMS Positive Test

Positive Criteria (Art. 2.3.2)	$ \Delta\delta^{13}\text{C} _{\text{ERC-TC}}$					
	T	E	A	Etio	5 α Adiol, 5 β Adiol	B, BM1, F 6 α -OH-AD, PS, PSL, EpiA *
i.	> 3.0 ‰				> 3.0 ‰ (either Adiol)	
ii.					> 3.0 ‰ (both Adioms)	
iii.		> 4.5 ‰				
iv.&			> 2.0 ‰	> 3.0 ‰		
v.			> 2.0 ‰		> 3.0 ‰ (either Adiol)	
				> 3.0 ‰	> 3.0 ‰ (either Adiol)	
vi.					$ \Delta\delta^{13}(\text{ERC-5}\alpha) > 4.0 \text{ ‰}$ and $\delta^{13}\text{C}(5\alpha) \leq -27.0 \text{ ‰}$	
vii.						> 4.0 ‰

* Determined from the steroid sulfoconjugated fraction.

& If T, 5 α Adiol and 5 β Adiol cannot be measured in the Sample.