

## TD2021NA version 2.0

### Summary of Major Modifications

The *Technical Document* on Harmonization of Analysis and Reporting of 19-Norsteroids related to Nandrolone, TD2021NA version 2.0, has been modified as follows:

#### Article 3.2.3

The requirement to inject a blank urine sample (19-NA below the Limit of Detection (LOD) of the method) before the *Sample* under confirmation has been removed. It is considered that the potential of any contamination shall be addressed by the Laboratory during method validation, and that the use of negative (NQC) and a positive quality control (PQC) samples is sufficient to monitor any issues during routine analysis.